

MNIST

Target Error Rate: 0.044	Number of test points: 17500	
Conjugate Prediction:	Error: 768.10 +- 30.04	Refuse: 294.40 +- 51.68
Isotonic Calibration:	Error: 35.70 +- 4.63	Refuse: 4680.10 +- 109.43
Sigmoid Calibration:	Error: 31.50 +- 5.04	Refuse: 4777.30 +- 93.77
Isotonic Calibration (Oracle):	Error: 766.00 +- 1.95	Refuse: 149.80 +- 36.45
Sigmoid Calibration (Oracle):	Error: 766.40 +- 2.01	Refuse: 143.10 +- 43.61
Target Error Rate: 0.022	Number of test points: 17500	
Conjugate Prediction:	Error: 342.10 +- 30.83	Refuse: 1957.00 +- 100.79
Isotonic Calibration:	Error: 17.30 +- 3.66	Refuse: 6059.60 +- 146.42
Sigmoid Calibration:	Error: 5.00 +- 2.32	Refuse: 8888.30 +- 271.77
Isotonic Calibration (Oracle):	Error: 359.20 +- 1.66	Refuse: 1202.20 +- 77.13
Sigmoid Calibration (Oracle):	Error: 359.70 +- 1.49	Refuse: 1179.80 +- 66.67
Target Error Rate: 0.011	Number of test points: 17500	
Conjugate Prediction:	Error: 159.40 +- 20.02	Refuse: 3087.70 +- 65.73
Isotonic Calibration:	Error: 9.80 +- 2.71	Refuse: 7356.40 +- 206.84
Sigmoid Calibration:	Error: 0.00 +- 0.00	Refuse: 17500.00 +- 0.00
Isotonic Calibration (Oracle):	Error: 168.00 +- 2.10	Refuse: 2201.00 +- 167.77
Sigmoid Calibration (Oracle):	Error: 168.60 +- 1.69	Refuse: 2163.50 +- 137.55

COD-RNA

Target Error Rate: 0.0316	Number of test points: 122142	
Conjugate Prediction:	Error: 3842.80 +- 111.59	Refuse: 617.00 +- 100.21
Isotonic Calibration:	Error: 342.70 +- 30.64	Refuse: 23305.80 +- 444.50
Sigmoid Calibration:	Error: 567.60 +- 34.31	Refuse: 18147.00 +- 259.48
Isotonic Calibration (Oracle):	Error: 3738.50 +- 86.64	Refuse: 843.30 +- 276.12
Sigmoid Calibration (Oracle):	Error: 3833.80 +- 6.87	Refuse: 766.80 +- 199.09
Target Error Rate: 0.0158	Number of test points: 122142	
Conjugate Prediction:	Error: 1819.90 +- 60.81	Refuse: 6923.40 +- 185.79
Isotonic Calibration:	Error: 196.50 +- 35.00	Refuse: 29241.30 +- 1474.63
Sigmoid Calibration:	Error: 160.60 +- 21.79	Refuse: 31358.90 +- 763.98
Isotonic Calibration (Oracle):	Error: 1773.10 +- 30.59	Refuse: 7160.30 +- 349.54
Sigmoid Calibration (Oracle):	Error: 1813.30 +- 5.00	Refuse: 7118.90 +- 282.96
Target Error Rate: 0.0079	Number of test points: 122142	
Conjugate Prediction:	Error: 853.10 +- 56.01	Refuse: 14336.80 +- 306.65
Isotonic Calibration:	Error: 85.10 +- 18.76	Refuse: 37939.90 +- 1712.54
Sigmoid Calibration:	Error: 10.20 +- 2.93	Refuse: 56771.90 +- 556.27
Isotonic Calibration (Oracle):	Error: 824.20 +- 19.81	Refuse: 14430.90 +- 471.01
Sigmoid Calibration (Oracle):	Error: 840.70 +- 9.97	Refuse: 14226.00 +- 587.18

SENSIT

Target Error Rate: 0.1665	Number of test points: 24632	
Conjugate Prediction:	Error: 4047.80 +- 118.37	Refuse: 299.60 +- 184.12
Isotonic Calibration:	Error: 1025.90 +- 67.41	Refuse: 9818.20 +- 224.58
Sigmoid Calibration:	Error: 1097.90 +- 49.24	Refuse: 9361.60 +- 157.07
Isotonic Calibration (Oracle):	Error: 4068.40 +- 18.50	Refuse: 181.60 +- 114.28
Sigmoid Calibration (Oracle):	Error: 4057.50 +- 18.21	Refuse: 252.50 +- 109.85
Target Error Rate: 0.0832	Number of test points: 24632	
Conjugate Prediction:	Error: 1385.80 +- 64.72	Refuse: 7865.90 +- 170.54
Isotonic Calibration:	Error: 155.00 +- 17.29	Refuse: 16627.30 +- 140.85
Sigmoid Calibration:	Error: 168.60 +- 15.60	Refuse: 16474.20 +- 123.94
Isotonic Calibration (Oracle):	Error: 1404.90 +- 26.95	Refuse: 7711.80 +- 305.72
Sigmoid Calibration (Oracle):	Error: 1400.60 +- 20.71	Refuse: 7767.80 +- 270.44
Target Error Rate: 0.0416	Number of test points: 24632	
Conjugate Prediction:	Error: 433.60 +- 22.91	Refuse: 14115.60 +- 210.21
Isotonic Calibration:	Error: 80.10 +- 11.12	Refuse: 17569.90 +- 176.92
Sigmoid Calibration:	Error: 3.10 +- 1.81	Refuse: 22122.70 +- 594.26
Isotonic Calibration (Oracle):	Error: 423.20 +- 15.50	Refuse: 14219.00 +- 218.64
Sigmoid Calibration (Oracle):	Error: 435.10 +- 3.36	Refuse: 14073.50 +- 79.18

SATI

Target Error Rate: 0.0757	Number of test points: 2718	
Conjugate Prediction:	Error: 194.30 +- 15.05	Refuse: 115.40 +- 32.47
Isotonic Calibration:	Error: 25.50 +- 5.57	Refuse: 769.50 +- 52.95
Sigmoid Calibration:	Error: 28.40 +- 5.54	Refuse: 749.40 +- 34.25
Isotonic Calibration (Oracle):	Error: 199.00 +- 2.68	Refuse: 70.70 +- 23.22
Sigmoid Calibration (Oracle):	Error: 199.60 +- 2.62	Refuse: 72.60 +- 27.66
Target Error Rate: 0.0378	Number of test points: 2718	
Conjugate Prediction:	Error: 90.80 +- 9.36	Refuse: 419.00 +- 32.38
Isotonic Calibration:	Error: 13.40 +- 2.97	Refuse: 945.50 +- 41.08
Sigmoid Calibration:	Error: 6.30 +- 3.41	Refuse: 1291.10 +- 132.45
Isotonic Calibration (Oracle):	Error: 88.90 +- 1.58	Refuse: 351.10 +- 38.69
Sigmoid Calibration (Oracle):	Error: 88.50 +- 1.43	Refuse: 357.70 +- 37.19
Target Error Rate: 0.0189	Number of test points: 2718	
Conjugate Prediction:	Error: 38.80 +- 8.57	Refuse: 675.40 +- 48.60
Isotonic Calibration:	Error: 6.40 +- 3.61	Refuse: 1148.90 +- 96.04
Sigmoid Calibration:	Error: 0.00 +- 0.00	Refuse: 2718.00 +- 0.00
Isotonic Calibration (Oracle):	Error: 38.10 +- 1.30	Refuse: 617.60 +- 51.73
Sigmoid Calibration (Oracle):	Error: 37.90 +- 2.02	Refuse: 635.70 +- 53.62

LETTER

Target Error Rate: 0.0448	Number of test points: 8750	
Conjugate Prediction:	Error: 363.10 +- 32.41	Refuse: 756.80 +- 163.37
Isotonic Calibration:	Error: 8.80 +- 3.22	Refuse: 3003.00 +- 93.42
Sigmoid Calibration:	Error: 0.70 +- 0.64	Refuse: 4150.70 +- 160.41
Isotonic Calibration (Oracle):	Error: 378.80 +- 2.09	Refuse: 236.70 +- 31.92
Sigmoid Calibration (Oracle):	Error: 377.50 +- 2.20	Refuse: 289.70 +- 48.99
Target Error Rate: 0.0224	Number of test points: 8750	
Conjugate Prediction:	Error: 157.90 +- 18.28	Refuse: 1894.60 +- 165.32
Isotonic Calibration:	Error: 4.60 +- 2.54	Refuse: 3516.10 +- 98.12
Sigmoid Calibration:	Error: 0.00 +- 0.00	Refuse: 8750.00 +- 0.00
Isotonic Calibration (Oracle):	Error: 174.60 +- 2.01	Refuse: 902.40 +- 71.03
Sigmoid Calibration (Oracle):	Error: 175.80 +- 1.94	Refuse: 866.50 +- 71.74
Target Error Rate: 0.0112	Number of test points: 8750	
Conjugate Prediction:	Error: 71.10 +- 7.27	Refuse: 2621.00 +- 101.90
Isotonic Calibration:	Error: 1.60 +- 1.20	Refuse: 4024.30 +- 79.82
Sigmoid Calibration:	Error: 0.00 +- 0.00	Refuse: 8750.00 +- 0.00
Isotonic Calibration (Oracle):	Error: 79.80 +- 1.17	Refuse: 1526.10 +- 81.53
Sigmoid Calibration (Oracle):	Error: 81.30 +- 0.64	Refuse: 1437.70 +- 47.77

CONNECT

Target Error Rate: 0.1636	Number of test points: 16890	
Conjugate Prediction:	Error: 2674.60 +- 59.20	Refuse: 555.80 +- 116.65
Isotonic Calibration:	Error: 470.40 +- 56.15	Refuse: 6954.80 +- 347.35
Sigmoid Calibration:	Error: 589.60 +- 30.69	Refuse: 6266.10 +- 76.41
Isotonic Calibration (Oracle):	Error: 2687.80 +- 63.20	Refuse: 101.80 +- 170.57
Sigmoid Calibration (Oracle):	Error: 2728.10 +- 26.35	Refuse: 11.60 +- 34.80
Target Error Rate: 0.0818	Number of test points: 16890	
Conjugate Prediction:	Error: 930.90 +- 35.76	Refuse: 5377.40 +- 177.43
Isotonic Calibration:	Error: 159.40 +- 23.03	Refuse: 9606.10 +- 272.42
Sigmoid Calibration:	Error: 191.10 +- 15.41	Refuse: 9204.00 +- 115.97
Isotonic Calibration (Oracle):	Error: 969.20 +- 32.54	Refuse: 4657.20 +- 216.17
Sigmoid Calibration (Oracle):	Error: 1010.60 +- 17.14	Refuse: 4495.80 +- 204.45
Target Error Rate: 0.0409	Number of test points: 16890	
Conjugate Prediction:	Error: 349.10 +- 14.56	Refuse: 8353.10 +- 149.44
Isotonic Calibration:	Error: 68.70 +- 20.76	Refuse: 11187.10 +- 360.78
Sigmoid Calibration:	Error: 45.90 +- 6.85	Refuse: 11751.90 +- 156.19
Isotonic Calibration (Oracle):	Error: 354.40 +- 25.75	Refuse: 7724.10 +- 292.64
Sigmoid Calibration (Oracle):	Error: 380.20 +- 9.35	Refuse: 7531.60 +- 219.04

COVER

Target Error Rate: 0.0854	Number of test points: 145253	
Conjugate Prediction:	Error: 11754.20 +- 217.51	Refuse: 6747.40 +- 574.90
Isotonic Calibration:	Error: 1643.20 +- 82.52	Refuse: 49448.10 +- 1001.77
Sigmoid Calibration:	Error: 2455.10 +- 55.73	Refuse: 41822.40 +- 356.16
Isotonic Calibration (Oracle):	Error: 12020.80 +- 66.08	Refuse: 4123.20 +- 319.41
Sigmoid Calibration (Oracle):	Error: 12037.20 +- 28.35	Refuse: 4185.70 +- 292.39
Target Error Rate: 0.0427	Number of test points: 145253	
Conjugate Prediction:	Error: 4838.70 +- 103.49	Refuse: 31047.10 +- 615.36
Isotonic Calibration:	Error: 700.20 +- 68.25	Refuse: 66739.50 +- 1657.28
Sigmoid Calibration:	Error: 898.50 +- 40.04	Refuse: 62854.60 +- 676.91
Isotonic Calibration (Oracle):	Error: 5125.20 +- 33.95	Refuse: 24802.30 +- 449.95
Sigmoid Calibration (Oracle):	Error: 5107.40 +- 22.14	Refuse: 25433.90 +- 443.10
Target Error Rate: 0.0214	Number of test points: 145253	
Conjugate Prediction:	Error: 1963.70 +- 65.74	Refuse: 52152.80 +- 730.40
Isotonic Calibration:	Error: 281.90 +- 19.08	Refuse: 83130.50 +- 1091.31
Sigmoid Calibration:	Error: 118.50 +- 14.02	Refuse: 97460.30 +- 1758.90
Isotonic Calibration (Oracle):	Error: 2152.30 +- 25.85	Refuse: 43909.10 +- 898.81
Sigmoid Calibration (Oracle):	Error: 2122.30 +- 20.78	Refuse: 45090.30 +- 638.69